



Measurement of Impedance Values of Different Erythrocyte Suspensions

Farklı Eritrosit Süspansiyonlarının Empedans Değerlerinin Ölçülmesi

İD Mehmet ÜYÜKLÜ

Bezmialem Vakıf University Faculty of Medicine, Department of Physiology, İstanbul, Turkey

ABSTRACT

Objective: The study was aimed to determine whether the impedance measurements of erythrocyte suspensions can be used with a different calculation method in determining the degree of erythrocyte aggregation.

Methods: Impedance measurements of different erythrocyte suspensions were recorded in horizontal glass capillaries with a haematocrit value of 40% with inductance-capacitance-resistance meter after a stop flow generated by the injector pump.

Results: As a result of calculating the impedance values of the samples for which the aggregation in phosphate-buffered saline didn't occur, the impedance values of the samples that aggregation at different grades took place were calculated as in the previous measurements of erythrocyte suspensions time course was significantly decreased for in diluted plasmas and was increased for in the 1% dextran 500 solution. It is obvious that aggregation index (AI) was calculated, using Z data exhibited a similar trend for the diluted plasma and plasma with 1% dextran 500, with a rank order of Dextran 500 > whole blood > 1/2 diluted plasma.

Conclusion: Although impedance measurements of erythrocyte suspensions don't allow for the calculation of erythrocyte aggregation kinetics, it is thought that they can be used as AI which indicates the degree of erythrocyte aggregation.

Keywords: Electrical measurements, impedance, capacitance, red blood cell aggregation

ÖZ

Amaç: Eritrosit süspansiyonları ile yapılan çalışmada ölçülen kapasitans değerlerinin değiştiği ve bu değişime de eritrosit agregasyonunun neden olduğu bildirilmiştir. Bu çalışmada, eritrosit süspansiyonlarının empedans ölçümlerinin eritrosit agregasyon derecesinin belirlenmesinde farklı bir hesaplama yöntemiyle kullanılıp kullanılamayacağının belirlenmesi amaçlanmaktadır.

Yöntemler: Farklı eritrosit süspansiyonlarının empedans ölçümleri hematokrit değerleri %40 olacak şekilde yatay cam kapillerde, enjektör pompasıyla oluşturulan beş saniyelik akım sonrasında endüktans-kapasitans-direnç metre ile kaydedilmiştir.

Bulgular: Fosfat tamponlu salin içinde agregasyonun meydana gelmediği örnekler ait empedans değerinin, farklı derecelerde agregasyonun meydana geldiği örnekler ait empedans değerlerinden çıkarılması sonucunda daha önceden yapılan ölçümlere benzer bir şekilde dekstran, tam kan ve dilüsyon örneklerinde benzer özellikte olmasına rağmen dilüsyon örneklerinde bir azalma, dekstran örneğinde ise artış olduğu görülmektedir. Empedans kayıtlarından hesaplanan agregasyon indeksleri (AI) büyükten küçüğe doğru dekstran 500 > tam kan > 1/2 dilüsyon şeklinde sıralanmaktadır ve bütün gruplar tam kan grubundan istatistiksel olarak farklıdır.

Sonuç: Eritrosit süspansiyonlarının empedans ölçümlerinin her ne kadar eritrosit agregasyon kinetiğinin hesaplanmasına imkan tanımasa da, eritrosit agregasyonunun derecesini belirten "agregasyon indeksi" olarak kullanılabileceği düşünülmektedir.

Anahtar Sözcükler: Elektriksel ölçüm, empedans, kapasitans, kırmızı kan hücresi toplanması

Address for Correspondence/Yazışma Adresi: Mehmet ÜYÜKLÜ, Bezmialem Vakıf University Faculty of Medicine, Department of Physiology, İstanbul, Turkey

E-mail: muyuklu@bezmialem.edu.tr **ORCID ID:** orcid.org/0000-0002-7100-9817

Received/Geliş Tarihi: 10.10.2018

Accepted/Kabul Tarihi: 29.12.2018

Cite this article as: Üyüklü M. Measurement of Impedance Values of Different Erythrocyte Suspensions. Bezmialem Science 2019;7(3):233-9.

©Copyright 2019 by the Bezmialem Vakıf University
Bezmialem Science published by Galenos Publishing House.

Introduction

Aggregation give rise to erythrocyte surfaces parallel to each other and clustering in a special way. Erythrocyte aggregation occurs only in the suspension phase in the presence of a certain size and structure of macromolecules. This clustering does not occur when the same cells suspended in simple salt solutions. Factors affecting red blood cell (RBC) aggregation can be divided into intrinsic and extrinsic factors: (1) Extrinsic factors include levels of plasma proteins [exempli gratia (e.g.), fibrinogen, macroglobulins], hematocrit and shear rate; (2) Intrinsic factors include RBC shape, deformability and membrane surface properties (1-3)

Aggregation can be altered during pathophysiological processes due to modifications of both plasma composition (e.g., in acute phase reactions) and cellular factors (e.g., increased oxidant stress) (3-5) RBC aggregation affects the flow properties of blood, especially at low shear rates, and therefore has the potential to influence blood flow in the circulatory system (6). The factors determining RBC aggregation could be classified in two groups, extracellular and cellular. The extracellular factors are: flow condition, concentration of plasma proteins (fibrinogen, immunoglobulins, albumin etc.), the presence of other macromolecules (e.g. dextrans), pH-value, osmotic pressure. The cellular factors are: the glycocalyx, determines kinetics, and extent of the contact area between the cells and so the rate and the strength of the aggregation. For example glutaraldehyde treatment of RBC results in intra- and intermolecular crosslinking membrane and cytosol constituents with drastically alteration of membrane viscoelastic and cytosol viscous properties with reduction in cell deformability and aggregation. Direct measurement of RBC aggregation has not been frequently used in routine clinical laboratories. The RBC aggregation characteristics can be analyzed using various methods and techniques which are described in the literature (7). Briefly, measurement of erythrocyte sedimentation rate (ESR) (8), low-shear viscometry (9), image analysis (10), dielectric analysis (11), ultrasound backscattering (12) electrical properties (13,14) and photometry (15) are among the methods which have been employed to quantify aggregation.

There are some reports assert that measuring the electrical properties of erythrocyte suspensions can be used for evaluation of erythrocyte aggregation (7,14,16,17). It was reported that electrical properties of erythrocyte suspensions showed a similar trend as photometric measurement and the parameters associated with aggregations could be calculated by using these records (13,16,18). In our previous study, we recorded the photometric and electrical properties of suspensions simultaneously during the erythrocyte aggregation through a glass capillary tube with a diameter of 800 μm . Recorded light transmittance-time curves (LT) showed a great similarity to the typical curves reflecting erythrocyte aggregation process (7,15,19,20).

This study was designed to compare the time course of electrical impedance and capacitance during RBC aggregation in a horizontal glass tube. RBC suspended in non-aggregating and

aggregating media were investigated; comparisons between measured modalities were based on calculated aggregation indexes that are accepted to reflect RBC aggregation (16) and are used by commercial instruments (15). In this study, it is aimed to determine whether the impedance measurements of erythrocyte suspensions can be used with a different calculation method in determining the degree of erythrocyte aggregation.

Methods

Blood Samples and Preparation of Red Blood Cell Suspensions

Venous blood samples were obtained from 10 healthy male volunteers, aged between 20 to 25 years, following the guidelines for hemorheological laboratory methods (16). The study was approved by the Clinical Research Ethical Committee of Bezmialem Vakıf University (19.09.2018). Informed consent was obtained from all participants, and study was conducted according to the Declaration of Helsinki Principles. In addition, adult male Wistar albino rats weighing between 250-300 g were used in this study. Blood samples were taken from abdominal aorta under ketamine-xylazine anesthesia from ten animals into sodium heparin (15 IU/mL). All animal use procedures were approved by the Laboratory Animals Ethical Committee, Bezmialem Vakıf University. Blood was centrifuged at 1400 g for 5 min, and plasma was separated carefully and saved. RBC were washed twice with isotonic phosphate-buffered saline (PBS, pH = 7.4). Samples were immediately evaluated, and studies on samples were completed within four hours following the blood drawing.

Modification of Red Blood Cell Aggregation: Red Blood Cell Suspended in Diluted Plasma

Plasma was diluted with PBS at ratios of 1/2 corresponding to 50% of the plasma components in undiluted plasma. Washed RBC from each donor were re-suspended either in autologous plasma (control) or in diluted plasma at 0.4 l/l hematocrit.

Modification of Red Blood Cell Aggregation: Red Blood Cell Suspended in 1% Dextran 500

Washed RBC from each donor were suspended in plasma containing 1% Dextran 500 (500 kDa, Sigma Chemical Company, St. Louis, MO, USA) at 0.4 l/l hematocrit.

Modification of Red Blood Cell Aggregation: Glutaraldehyde Treatment of Red Blood Cell

Washed RBC from each donor were re-suspended in PBS at 0.05 l/l hematocrit, following which glutaraldehyde (GA, Merck, Darmstadt, Germany) was added to achieve a final concentration of 0.003% and the suspension incubated at room temperature (22 °C) for 30 min. RBC were then washed three times with PBS after the incubation and re-suspended in autologous plasma at 0.4 l/l hematocrit.

Measurement System and Experimental Protocol

RBC aggregation in glass capillaries with internal diameters of 800 μm was assessed by monitoring electrical impedance

during aggregation process. The measurement system is outlined in Figure 1. Electrical impedance of RBC suspensions were measured in horizontally-aligned, 75 mm long capillaries, across the two stainless steel electrodes attached to both ends of the capillary. The measurement chamber (glass capillary and electrodes) were placed in a temperature-controlled box and maintained at 37 °C. A computer controlled inductance-capacitance-resistance (LCR)-meter (Hioki, 3532 LCR HiTester, Nagano, Japan) was used for continuous monitoring of electrical impedance parameters between these electrodes, during the RBC aggregation process. The instrument was operated in series mode and a two-electrode configuration was applied. Briefly, the LCR meter applies a test signal across the electrodes at 100 kHz and 1 volt constant voltage (V), measures the resulting current (I) and phase angle (θ) between V and I.

RBC suspensions were pumped through the capillary with a syringe pump (World Precision Instruments, Aladdin-4000, Florida, USA) at a calculated flow rate to generate a wall shear rate of $\sim 500 \text{ s}^{-1}$, for 5 s, to obtain complete disaggregation, at the start of each measurement procedure. Electrical impedance parameters were recorded at 1 s intervals for the 120 s following the stoppage of flow in the capillary tube, via a RS-232 interface on a digital computer.




Figure 1. Measurement system
LCR: Inductance-capacitance-resistance




Figure 2. Calculation procedure for aggregation parameters, using light transmittance -time data. "A" and "B" are the areas of the surfaces above and below the curve, respectively. Aggregation index is equal to $B / (A+B)$
AI: Aggregation index, LT: Light transmittance

Calculations and Statistics

It is known that when erythrocyte aggregation is examined by its property of light reflection or transmittance, the course of time can be modelled by a curve with double-exponential Time courses of C and Z data recorded during the aggregation process after abruptly stopping flow were used to calculate aggregation index (AI) which is equivalent to those reported by commercially-available RBC aggregometers (Figure 2). Calculations were conducted using a special software written in LabView 8.6 (National Instruments, Austin, Texas, USA). The results of the measurements on RBC suspensions with normal and modified aggregation are expressed as mean $\pm$ standard error. Statistical comparisons were done using one-way analysis of variance followed with Bonferroni's multiple comparison test.

Results

Time Course of C and Z During Red Blood Cell Aggregation

Figure 3a demonstrates the time course C, recorded using the same RBC suspensions prepared in 1/2 diluted plasma, dextran and whole plasma. There is a sharp decrease in C following the stoppage of flow, reaching to a minimum in a few second. This valley is followed by increased C with a characteristic time course. The initial phase of decreased C reflects the recovery of RBC morphology for cells initially deformed by shear forces during high-shear flow (corresponding to $\sim 500 \text{ s}^{-1}$ shear rate), with the shape changing from an elongated form to the normal biconcave-discoid shape. The later phase of gradual increases of C reflects RBC aggregate formation which follows a characteristic time course.




Figure 3. (a) Time course of series capacitance (C) and (b) Impedance (Z) across the glass capillary after a sudden stop of flow of red blood cell suspended in dextran, whole blood and 1/2 diluted plasma. Capacitance and impedance were monitored at 100 KHz frequency
PBS: Phosphate-buffered saline

Capacitance recorded in RBC suspension in dextran, whole blood and 1/2 diluted plasma exhibited similar pattern, although the time course was significantly slower in 1/2 diluted plasma. In capacitance measurements, the changes of plasma samples diluted by PBS after the flow stopped showed a little peak than the plasma samples not diluted by PBS (Figure 3a). There are significant changes in the capacitance measurements in all samples.

The time course of the change in Z after reaching to the peak values were similar to each other for RBC suspensions in PBS, diluted and undiluted plasma (Figure 3b). A finding of this study was the time course of Z changes for non-aggregating suspensions of RBC in PBS. Z changes had the same pattern as those for aggregating suspensions, while such similarity between aggregating and non-aggregating suspensions was not observed for C.

Aggregation Characteristics of Red Blood Cell Suspensions

RBC aggregation parameters calculated using syllectogram recorded from different RBC suspensions were compared by analyzing data obtained using samples with whole blood, decreased (i.e., 1/2 diluted plasma), and enhanced (i.e., 1% dextran 500) aggregation. Aggregation parameters for these samples, as measured by a widely used commercial erythrocyte aggregometer [i.e., Laser-assisted optical rotational cell analyzer (LORCA)], indicated the expected differences (Table 1) (14,18). The AI parameter was significantly decreased for RBC suspended in 1/2 diluted plasma and was increased for RBC suspended in the 1% Dextran 500 solution. It is obvious that AI calculated using C data exhibited a similar trend for the diluted plasma and plasma with 1% dextran 500, with a rank order of plasma + D500 >undiluted plasma >1/2 diluted plasma. This trend was not seen when using Z data.

Electrical Characteristics of Red Blood Cell -plasma Suspensions and Red Blood Cell Aggregation Indexes Calculated Using Z Time Course

In Figure 4, it is seen that the impedance values of the blood samples prepared in the second minute after stopping the flow are increased. Impedance values in PBS and 40% hematocrit adjusted samples in plasma showed statistically significant increases in both PBS and plasma samples without erythrocytes

($p<0.001$). An interesting result is that although the hematocrit ratios are the same, the impedance value of erythrocyte suspension prepared in plasma is significantly different from the PBS sample ($p<0.001$). However, there was no significant difference between PBS and plasma in which there was no cell ($p>0.05$).

In Figure 5, the impedance values of the erythrocyte suspension in the plasma are increased due to the increase of the hematocrit value. Gradually increasing the hematocrit values from 20% to 40% to 60% by addition erythrocyte into the plasma significantly increases the impedance values ($p<0.001$). As shown in Figure 4 and Figure 5, there is a significant difference between the




Figure 4. Impedance values of phosphate-buffered saline and plasma samples adjusted to 40% hematocrit. Data are mean $\pm$ standard error (n=10). Difference from "phosphate-buffered saline and plasma"; ***, $p<0.001$, +, $p<0.05$

PBS: Phosphate-buffered saline




Figure 5. Effect of hematocrit ratio on impedance value. Data are mean $\pm$ standard error (n=10). Difference from "phosphate-buffered saline", ***, $p<0.001$. Difference from "Plasma", +, $p<0.05$, +, $p<0.01$, +, $p<0.001$

PBS: Phosphate-buffered saline

Table 1. Aggregation Index values by the commercial erythrocyte aggregometer (i.e., Laser-assisted optical rotational cell analyzer, Myrenne Aggregometer) (14,18) Impedance and Capacitance for red blood cell suspensions prepared in autologous plasma, 1/2 diluted autologous plasma with phosphate-buffered saline corresponding to the 50% of plasma components and autologous plasma with 1% dextran 500 in. Data are mean $\pm$ standard error (n=10)

Aggregation index (AI) (au)	LORCA (AI)	Myrenne (M index)	Capacitance (C)	Impedance (Z)
Whole blood	70.3 $\pm$ 2.3	19.9 $\pm$ 1.1	79.0 $\pm$ 1.7	73.8 $\pm$ 2.5
1/2 diluted	32.6 $\pm$ 2.4***	2.5 $\pm$ 1.0**	60.6 $\pm$ 5.1**	59.6 $\pm$ 1.5**
1% 500 kDa Dextran	82.1 $\pm$ 1.5*	33.2 $\pm$ 1.6**	84.4 $\pm$ 1.8*	69.8 $\pm$ 5.3
PBS	ND	ND	ND	34.5 $\pm$ 2.3**

Difference from "whole blood"; *, $p<0.05$; **, $p<0.01$; ***, $p<0.001$. au: arbitrary unit ND: Not determined, AI: Aggregation index, PBS: Phosphate-buffered saline, LORCA: Laser-assisted optical rotational cell analyzer

impedance values of erythrocyte suspensions in PBS and plasma with the same hematocrit value (40%). This shows us that the value of the impedance changes according to both the properties of the suspension environment and the number of cells it contains. The impedance values of samples with 40% hematocrit




Figure 6. Comparison of phosphate-buffered saline and plasma impedance values. Data are mean $\pm$ standard error (n=10). Difference from "phosphate-buffered saline, 0.9% NaCl and plasma", ***: $p < 0.001$, +++: $p < 0.001$
PBS: Phosphate-buffered saline, RBC: Red blood cell




Figure 7. Comparison of calculated impedance values for (a) Human and (b) Rat erythrocyte suspensions. Data are mean $\pm$ standard error (n=10). Difference from "whole blood"; *: $p < 0.05$; **: $p < 0.01$, ***: $p < 0.001$
AI: Aggregation index




Figure 8. (a-b) Impedance and (c) Aggregation index values of whole blood and GA groups. Data are mean $\pm$ standard error (n=10). Difference from "whole blood"; *: $p < 0.05$; ***: $p < 0.001$
AI: Aggregation index, GA: Glutoraldehyde

prepared in PBS and the impedance values of samples with 20% hematocrit in plasma shown in Figure 5 are very close to each other. This may also be related to the behavior of erythrocytes, besides the number of cells and suspension medium. It is known that erythrocyte aggregation does not occur in simple salt solutions in which there are no macromolecules. Higher values of impedance in samples prepared in plasma may be associated with erythrocyte aggregation. The rise in the impedance value with the increase in hematocrit also explains this case.

In simple salt solutions (eg PBS and 0.9% NaCl) in which no macromolecule is present, rouleaux formation does not occur. As shown in Figure 6, the impedance values of PBS and 0.9% NaCl are very close to each other. Erythrocytes in suspension are responsible for the increased impedance in 40% erythrocyte suspension prepared in PBS. When the PBS and/or 0.9% NaCl values are compared with the plasma, the impedance value of the plasma is higher. When the PBS and/or 0.9% NaCl values are compared with the plasma, the impedance value of the plasma is higher. Although this difference is not significant ($p > 0.05$), it is thought that the macromolecules in the plasma may be responsible for this increment (eg, fibrinogen, etc.). It is assumed that this significant difference ($p < 0.001$) between the two samples that only the suspension medium is different with the same hematocrit represents erythrocyte aggregation.

Figure 7 shows that when it presented in terms of kilo ohm there is a significant difference ($p < 0.001$) between the impedance value of the erythrocyte suspension in the hematocrit adjusted to 40% and the impedance value of the erythrocyte suspension in the plasma with the hematocrit of 40%. Figure 7a shows that the value of dilution prepared with human blood decreased significantly ($p < 0.01$) compared to the whole blood group and increased in the dextran group ($p < 0.05$). When the same study was repeated with rat blood (Figure 7b), the same changes were observed. In dilution and dextran samples where aggregation levels have been modified, a similar trend is seen with photometric devices commonly used for measuring erythrocyte aggregation (Dextran>Whole Blood>Dilution).

Figures 8a and 8b show impedance values for whole blood and GA groups. The impedance value of the erythrocytes treated with GA and suspension prepared with 40% hematocrit with its own plasma was found to be lower than the whole blood impedance value ($p < 0.05$). These results show that in erythrocytes prepared with the same plasma with the same hematocrit ratio, only the GA-induced structural change resulted in reduced measured impedance values. Similarly, the calculated AI value was significantly reduced compared to the whole blood group (Figure 8c, $p < 0.001$).

Discussion

Aggregation denotes the formation of reversible clumps of RBC under sufficiently low shear stresses (21,22). It has been reported that electrical properties of red cell suspensions also change during aggregation (11,13,14,23-27). The electrical properties of blood plasma and blood cells differ from each other. The

plasma and cell interior consist of conducting fluids with certain electrical resistivities and can be simulated by resistors, whereas the cell membranes consist of phospholipids and proteins with dielectric properties and can be simulated by capacitors. Hence, the electrical impedance of blood is primarily characterized by three parameters: plasma resistance R_p , cell interior resistance R_i and cell membrane capacitance C (28).

Electrical properties of erythrocyte suspensions are determined by various factors. They include the suspension characteristics (suspension medium, hematocrit and RBC properties), the hydrodynamic conditions (shear forces), the measurement conditions (geometry of the measuring system, the electrodes and measuring the frequency characteristics). Prior studies have suggested that electrical properties of the suspensions may reflect the kinetics of RBC aggregation if monitored during the aggregation process.

In our previous studies, C recordings were also obtained using RBC suspended in isotonic PBS, a suspending medium which did not contain proteins or macromolecules. RBC do not aggregate if they are suspended in such media, and this lack of aggregation was confirmed by “zero” aggregation indexes provided by the photometric aggregometer. The time course of C after the sudden stop exhibited a different pattern than Z . There was no detectable change of C for the nonaggregating RBC suspension in PBS. An expected finding of this study was the time course of Z changes for nonaggregating suspensions of RBC in PBS (Table 1). Z changes had the same pattern as those for aggregating suspensions, while such similarity between aggregating and nonaggregating suspensions was not observed for C .

Impedance values in PBS and 40% hematocrit adjusted samples in plasma showed statistically significant increases in both PBS and plasma samples without erythrocytes (Figure 4). An interesting result is that although the hematocrit ratios are the same, the impedance value of erythrocyte suspension prepared in plasma is significantly different from the PBS sample ($p < 0.001$). However, there was no significant difference between PBS and plasma in which there was no cell ($p > 0.05$).

It was reported that measuring the electrical properties of erythrocyte suspensions can be one of the important methods which evaluate the erythrocyte aggregation (11,25-27). In these reports, the electrical properties of plasma and erythrocyte suspensions were evaluated by impedance and dielectric spectroscopy (11,19,26,29). Zhao and Jacopson (29) showed that the values of electrical impedance and capacitance were increased by fibrinogen concentration in suspensions. They also report that ESR was correlated with capacitance and capacitance was very sensitive to erythrocyte sedimentation rate. Pribush et al. (11,26,27) demonstrated that electrical properties of erythrocyte suspensions changed during aggregation. They speculated that monitoring of the electrical properties of these suspensions could give information about the process of aggregations. Capacitance values measured after the sudden cessation of blood flow in the system has changed and this change caused by the RBC aggregation were reported in studies using whole blood and

erythrocyte suspensions (11). Electrical properties of erythrocyte suspensions showed a similar trend as photometric measurement and the parameters associated with aggregations could be calculated by using these records were reported (13,14,25).

The LORCA records reflected light (LR) from RBC suspensions while the Myrenne Aggregometer record transmitted light (LT). The three types of RBC suspensions (i.e., suspensions in 1/2 diluted plasma, whole blood and plasma containing 1% dextran 500) exhibited different aggregation characteristics, judged by LORCA and Myrenne aggregometers (Table 1). RBC suspended in 1/2 diluted plasma (50% of original plasma components) represent decreased aggregation, while RBC suspended in 1% dextran 500 represent markedly increased aggregation; both changes of aggregation are due to alterations of suspending phase composition. The decreased concentrations of plasma components due to dilution resulted in significant and progressive changes of AI as measured by the three methods (LT, LR and C).

In our previous studies (13,17), we recorded the photometric and electrical properties of suspensions simultaneously during the erythrocyte aggregation of glass capillary tubes with a diameter of 1000 μm . We evaluated the correlations of AI of erythrocyte suspensions measured by photometric aggregometer and AI obtained from calculating the area under curve of light transmission, impedance and capacitance measurements and found that there was a positive and statistically significant correlation between capacitance value/light transmission and AI. There was no significantly correlation between impedance and AI value. According to results of this pre study, it was suggested that capacitance values monitored during erythrocyte aggregation in capillary tubes were reflected the aggregation time period. This suggestion was stated by comparisons using simultaneously records of LT.

When the PBS and/or 0.9% NaCl values are compared with the plasma, the impedance value of the plasma is higher. Although this difference is not significant ($p > 0.05$), it is thought that the macromolecules in the plasma may be responsible for this increment (eg, fibrinogen, etc.). It is assumed that this significant difference ($p < 0.001$) between the two samples that only the suspension medium is different with the same hematocrit represents erythrocyte aggregation (Figure 6 and 7). Only, the GA-induced structural change resulted in reduced measured impedance values in erythrocytes prepared with the same plasma with the same hematocrit ratio. Similarly, the calculated AI value was significantly reduced compared to the whole blood group (Figure 8). It was obvious from all these results, the impedance value of the erythrocyte suspensions is influenced by the composition of the suspension medium, the number of erythrocytes in suspension (hematocrit) and the structural properties of erythrocytes.

Conclusions

We aimed in this study to determine whether the impedance measurements of erythrocyte suspensions can be used with a different calculation method in determining the degree of

erythrocyte aggregation. Although impedance measurements of erythrocyte suspensions do not allow for the calculation of erythrocyte aggregation kinetics (T fast and $t_{1/2}$), it is thought that they can be used as “AI” which indicates the degree of erythrocyte aggregation. In this way, AI can be measured in a simpler and repeatable way.

Ethics

Ethics Committee Approval: The study was approved by the Clinical Research Ethical Committee of Bezmialem Vakıf University (19.09.2018).

Informed Consent: Informed consent was obtained from all participants, and study was conducted according to the Declaration of Helsinki Principles.

Peer-review: Externally peer-reviewed.

References

1. Baskurt OK. In vivo correlates of altered blood rheology. *Biorheology* 2008;45:629-38.
2. Meiselman HJ. Red blood cell aggregation: 45 years being curious. *Biorheology* 2009;46:1-19.
3. Rampling MW, Meiselman HJ, Neu B, Baskurt OK. Influence of cell-specific factors on red blood cell aggregation. *Biorheology* 2004;41:91-112.
4. Rampling MW. Haemorheology and the inflammatory process. *Clin Hemorheol Microcirc* 1998;19:129-32.
5. Baskurt OK, Temiz A, Meiselman HJ. Red blood cell aggregation in experimental sepsis. *J Lab Clin Med* 1997;130:183-90.
6. O.K.Baskurt, M.R. Hardeman, M.W. Rampling and H.J. Meiselman. In vivo hemorheology iHoHaH, IOS Press, Amsterdam 2007;322-38.
7. Baskurt OK, Meiselman HJ, Kayar E. Measurement of red blood cell aggregation in a „plate-plate“ shearing system by analysis of light transmission. *Clin Hemorheol Microcirc* 1998;19:307-14.
8. Houbouyan LL, Delamaire M, Beauchet A, Gentil M, Cauchois G, Taccoen A, et al. Multicenter study of an erythro-aggregometer: quality control and standardization. *Clin Hemorheol Microcirc* 1997;17:299-306.
9. Baskurt OK, Meiselman HJ. Cellular determinants of low-shear blood viscosity. *Biorheology* 1997;34:235-47.
10. Berliner S, Zeltser D, Shapira I, Assayag EB, Mardi T, Serov J, et al. A simple biomarker to exclude the presence of low grade inflammation in apparently healthy individuals. *J Cardiovasc Risk* 2002;9:281-6.
11. Pribush A, Meiselman HJ, Meyerstein D, Meyerstein N. Dielectric approach to the investigation of erythrocyte aggregation: I. Experimental basis of the method. *Biorheology* 1999;36:411-23.
12. Rouffiac V, Peronneau P, Hadengue A, Barbet A, Delouche P, Dantan P, et al. A new ultrasound principle for characterizing erythrocyte aggregation: in vitro reproducibility and validation. *Invest Radiol* 2002;37:413-20.
13. Baskurt OK, Uyuklu M, Meiselman HJ. Simultaneous monitoring of electrical conductance and light transmittance during red blood cell aggregation. *Biorheology* 2009;46:239-49.
14. Baskurt OK, Uyuklu M, Meiselman HJ. Time course of electrical impedance during red blood cell aggregation in a glass tube: comparison with light transmittance. *IEEE Trans Biomed Eng* 2010;57:969-78.
15. Hardeman MR, Dobbe JG, Ince C. The Laser-assisted Optical Rotational Cell Analyzer (LORCA) as red blood cell aggregometer. *Clin Hemorheol Microcirc* 2001;25:1-11.
16. Baskurt OK, Boynard M, Cokelet GC, Connes P, Cooke BM, Forconi S, et al. New guidelines for hemorheological laboratory techniques. *Clin Hemorheol Microcirc* 2009;42:75-97.
17. Baskurt OK, Uyuklu M, Hardeman MR, Meiselman HJ. Photometric measurements of red blood cell aggregation: light transmission versus light reflectance. *J Biomed Opt* 2009;14:540-44.
18. Baskurt OK, Uyuklu M, Ulker P, Cengiz M, Nemeth N, Alexy T, et al. Comparison of three instruments for measuring red blood cell aggregation. *Clin Hemorheol Microcirc* 2009;43:283-98.
19. Gaw RL, Cornish BH, Thomas BJ. The electrical impedance of pulsatile blood flowing through rigid tubes: a theoretical investigation. *IEEE Trans Biomed Eng* 2008;55:721-7.
20. Kiesewetter H, Radtke H, Schneider R, Mussler K, Scheffler A, Schmid-Schonbein H. [The mini-erythrocyte aggregometer: a new apparatus for the rapid quantification of the extent of erythrocyte aggregation]. *Biomed Tech (Berl)* 1982;27:209-13.
21. M.W. Rampling. Red cell aggregation and yield stress. In: *Clinical blood rheology*; Lowe GDO ECP, Inc Florida 1988;45-64.
22. H.J. Meiselman. Red blood cell role in RBC aggregation: 1963-1993 and beyond. *Clin Hemorheol Microcirc* 1993;13:575-92.
23. Antonova N, Riha P, Ivanov I. Time dependent variation of human blood conductivity as a method for an estimation of RBC aggregation. *Clin Hemorheol Microcirc* 2008;39:69-78.
24. Balan C, Balut C, Gheorghe L, Gheorghe C, Gheorghiu E, Ursu G. Experimental determination of blood permittivity and conductivity in simple shear flow. *Clin Hemorheol Microcirc* 2004;30:359-64.
25. Pribush A, Hatzkelson L, Meyerstein D, Meyerstein N. A novel technique for quantification of erythrocyte aggregation abnormalities in pathophysiological situations. *Clin Hemorheol Microcirc* 2007;36:121-32.
26. Pribush A, Meiselman HJ, Meyerstein D, Meyerstein N. Dielectric approach to investigation of erythrocyte aggregation. II. Kinetics of erythrocyte aggregation-disaggregation in quiescent and flowing blood. *Biorheology* 2000;37:429-41.
27. Pribush A, Meyerstein D, Meyerstein N. Conductometric study of shear-dependent processes in red cell suspensions. I. Effect of red blood cell aggregate morphology on blood conductance. *Biorheology* 2004;41:13-28.
28. Fricke H, Morse S. The Electric Resistance and Capacity of Blood for Frequencies between 800 and 4(1/2) Million Cycles. *J Gen Physiol* 1925;9:153-67.
29. Zhao TX, Jacobson B. Quantitative correlations among fibrinogen concentration, sedimentation rate and electrical impedance of blood. *Med Biol Eng Comput* 1997;35:181-5.